



OFFICIAL LISTING

NSF International Certifies that the products appearing on this Listing conform to the requirements of NSF/ANSI Standard 49 - Biosafety Cabinetry: Design, Construction, Performance, and Field Certification

This is the Official Listing recorded on July 31, 2014.

Esco Micro Pte Ltd
21 Changi South Street 1
486777
Singapore
65 6542 0833



Facility: Bintan Island, Indonesia

Model Number	Cabinet Type/Style		Inflow Velocity (fpm)	CBV (cfm)		Cabinet Width ft.	Bench Window Ht		Acceptable Options
				Downflow Velocity (fpm)	at Static Pressure (in w.g.)		Ht/Type	Max in. in.	
AC2-4N7 ^[5] ^[6]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
AC2-4S8-NS ^[2] ^[3] ^[5]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
AC2-4S9-NS ^[5]	A2	A		55-65	N/A	4	8S 36	I.V. Pole U.V. Light	
AC2-4Y7 ^[5] ^[6]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
AC2-6N7 ^[6] ^[7]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
AC2-6S8-NS ^[2] ^[3] ^[7]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
AC2-6S9-NS ^[7]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
AC2-6Y7 ^[6] ^[7]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
AR2-4S9 ^[5]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
AR2-6S9 ^[7]	A2	A		55-65	N/A		8S 36	I.V. Pole U.V. Light	
LA2-4A1-E ^[1] ^[2]	A2	A	100-110	65-75	N/A	4	9S 34	I.V. Pole U.V. Light	
LA2-4A2-E ^[1]	A2	A	100-110	65-75	N/A	4	9S 34	I.V. Pole U.V. Light	
LA2-4A3-E ^[1] ^[3]	A2	A	100-110	65-75	N/A	4	9S 34	I.V. Pole U.V. Light	
LA2-5A1-E ^[1] ^[2]	A2	A	100-110	65-75	N/A	5	9S 36	I.V. Pole U.V. Light	
LA2-5A2-E ^[1]	A2	A	100-110	65-75	N/A	5	9S 36	I.V. Pole U.V. Light	
LA2-5A3-E ^[1] ^[3]	A2	A	100-110	65-75	N/A	5	9S 36	I.V. Pole U.V. Light	
LA2-6A1-E ^[1] ^[2]	A2	A	100-110	60-70	N/A	6	8S 35	I.V. Pole U.V. Light	
LA2-6A2-E ^[1]	A2	A	100-110	60-70	N/A	6	8S 35	I.V. Pole U.V. Light	

Note: Additions shall not be made to this document without prior evaluation and acceptance by NSF International.



LA2-6A3-E ^[1] [3]	A2	A	100-110	60-70	N/A	6	8S 35	I.V. Pole U.V. Light
LB2-4B1-E ^[2] [4]	B2	A	100-110	55-65	803 @ 1.9	4	8S 36	I.V. Pole U.V. Light
LB2-4B2-E ^[4]	B2	A	100-110	55-65	803 @ 1.9	4	8S 36	I.V. Pole U.V. Light
LB2-4B3-E ^[3] [4]	B2	A	100-110	55-65	803 @ 1.9	4	8S 36	I.V. Pole U.V. Light
LB2-6B1-E ^[2] [4]	B2	A	100-110	55-65	1269 @ 2.1	6	8S 36	I.V. Pole U.V. Light
LB2-6B2-E ^[4]	B2	A	100-110	55-65	1269 @ 2.1	6	8S 36	I.V. Pole U.V. Light
LB2-6B3-E ^[3] [4]	B2	A	100-110	55-65	1269 @ 2.1	6	8S 36	I.V. Pole U.V. Light
LR2-4S1-E ^[1] [2]	A2	A	100-110	65-75	N/A	4	9S 34	I.V. Pole U.V. Light
LR2-4S2-E ^[1]	A2	A	100-110	65-75	N/A	4	9S 34	I.V. Pole U.V. Light
LR2-4S3-E ^[1] [3]	A2	A	100-110	65-75	N/A	4	9S 34	I.V. Pole U.V. Light
LR2-5S1-E ^[1] [2]	A2	A	100-110	65-75	N/A	5	9S 36	I.V. Pole U.V. Light
LR2-5S2-E ^[1]	A2	A	100-110	65-75	N/A	5	9S 36	I.V. Pole U.V. Light
LR2-5S3-E ^[1] [3]	A2	A	100-110	65-75	N/A	5	9S 36	I.V. Pole U.V. Light
LR2-6S1-E ^[1] [2]	A2	A	100-110	60-70	N/A	6	8S 35	I.V. Pole U.V. Light
LR2-6S2-E ^[1]	A2	A	100-110	60-70	N/A	6	8S 35	I.V. Pole U.V. Light
LR2-6S3-E ^[1] [3]	A2	A	100-110	60-70	N/A	6	8S 35	I.V. Pole U.V. Light

[1] Inflow nominal set point of 105 fpm was established with a direct airflow reading instrument. This nominal set point was confirmed using the manufacturer's recommended alternate method with thermal anemometer in a constricted (3 inch high) access opening (consult manufacturer for appropriate correction factor, if applicable), without adjusting cabinet airflow balance. Downflow nominal set point of 70 fpm was established with I.V. Pole and U.V. Light removed. This cabinet model was Certified to NSF/ANSI 49-2011.

[2] Certified for use with a power supply of 230V/50Hz.

[3] Certified for use with a power supply of 230V/60Hz.

[4] Inflow nominal set point of 105 fpm was established with a direct airflow reading instrument. This nominal set point was confirmed using the manufacturer's recommended alternate method with thermal anemometer in a constricted (3 inch high) access opening (consult manufacturer for appropriate correction factor, if applicable), without adjusting cabinet airflow balance. Downflow nominal set point of 60 fpm was established with I.V. Pole and U.V. Light removed. This cabinet model was Certified to NSF/ANSI 49-2012.

[5] Beginning with serial number 2014 - 89128. Inflow nominal set point of 105 fpm was established with a direct airflow reading instrument. This nominal set point was confirmed using the manufacturer's recommended alternate method with thermal anemometer in a constricted (3 inch high) access opening (consult manufacturer for appropriate correction factor, if applicable), without adjusting cabinet airflow balance. Downflow nominal set point of 60 fpm was established with I.V. Pole and U.V. Light removed. This cabinet model was Certified to NSF/ANSI 49-2012.

[6] Approved for alternate power modes of 100V/50Hz and 100V/60Hz.

[7] Beginning with serial number 2014 - 84641. Inflow nominal set point of 105 fpm was established with a direct airflow reading instrument. This nominal set point was confirmed using the manufacturer's recommended alternate method with thermal anemometer in a constricted (3 inch high) access opening (consult manufacturer for appropriate correction factor, if applicable), without adjusting cabinet airflow balance. Downflow nominal set point of 60 fpm was established with I.V. Pole and U.V. Light removed. This cabinet model was Certified to NSF/ANSI 49-2012.

Note: Additions shall not be made to this document without prior evaluation and acceptance by NSF International.